## **Features**

# Regulated Converter

- Medical certified 2MOPP module, BF ready
- Class II installations (without FG)
- IP68 waterproof encapsulation
- Operation altitude certified up to 5000m
- No external components necessary
- Energy Efficiency Level VI

#### **Description**

The RACM18-ER series comprises highly reliable power conversion modules in a potted IP68 waterproof encapsulation to withstand harsh operating conditions. With a certified operation up to 5000m altitude and a temperature range from -20°C up to +80°C these modules are built to power medical healthcare, household, sanitary, smart building and automation process appliances. For easy integration the product line is certified to medical, household, as well as to ITE safety standards and designed to pass class B limits by more than 6dB margin without the need for any external components.


Selection Guide				
Part Number	Input Voltage Range [VAC]	Output Voltage <sup>(1)</sup> [VDC]	Output Current [A]	Efficiency typ. <sup>(2)</sup> [%]
RACM18-05SER (3)	90-264	5	2.5	81
RACM18-12SER (3)	90-264	12	1.5	82
RACM18-24SER (3)	90-264	24	0.75	83.5

#### Notes:

Note1: Other output voltages on request

Note2: Efficiency is tested at nominal input (115/230VAC) and full load at +25°C ambient

## **Model Numbering**



Notes:

Note3: Other connection types on request



## RACM18-ER

# 18 Watt Round Shape Single Output






















IEC/EN60950-1 certified
UL60950-1 certified
ANSI/AAMI ES60601-1 certified
IEC/EN60601-1 certified
IEC/EN60335-1 certified
IEC/EN61558-1 certified
IEC/EN61558-2-16 certified
IEC/EN60601-1-2 certified
EN55024/32 certified
EN55014-1 (-2) certified
CISPR32 certified





## **Series**

## Specifications (measured @ ta= 25°C, nom. Vin (115/230VAC), full load after warm-up unless otherwise stated)

BASIC CHARACTERISTICS				
Parameter	Condition	Min.	Тур.	Max.
Internal Input Filter				Pi type
Input Voltage Range		90VAC	230VAC	264VAC
Input Current	115VAC 230VAC			500mA 150mA
Inrush Current	115VAC 230VAC		24A 46A	
No load Power Consumption			40mW	75mW
Input Frequency Range		47Hz		63Hz
Minimum Load		0%		
Power Factor			0.46	
Start-up Time	115VAC 230VAC		180ms 200ms	
Rise Time	115VAC/230VAC		15ms	
Hold-up Time	115VAC 230VAC		15ms 65ms	
Internal Operating Frequency	100% load at nominal Vin		100kHz	
Output Ripple and Noise	20MHz BW			140mVp-p








REGULATIONS		
Parameter	Condition	Value
Output Accuracy		±3.0% max.
Line Regulation	low line to high line	±1.0% max.
Load Regulation	0% to 100% load	±1.0% max.
Transient Response	100% load step change	±3.0% max.
	continued on next page	



## **Series**

## Specifications (measured @ ta= 25°C, nom. Vin (115/230VAC), full load after warm-up unless otherwise stated)




Parameter	Туре			Value
Input Fuse		line & neutral)	T2A, slow bl	
Short Circuit Protection (SCP)	momar	into a noutraly	continuo	us, auto recovery
onort official Frotection (oor)	5\/o	ıt, 12Vout		6VDC, Latch OFF
Over Voltage Protection (OVP)		24Vout		4VDC, Latch OFF
Over Voltage Category (OVC)			_	OVCII
		90VAC	145% of nominal Output Current, auto recovery	
Over Current Protection (OCP)	< 1 minute	160VAC	180% of nominal Output Current, auto recovery	Hiccup Mode
		264VAC	165% of nominal Output current, auto recovery	
Over Temperature Protection (OTP)	95°C ambient		thermal shutdov	vn, auto recovery
Class of Equipment				Class II
Isolation Voltage (3)	I/P to O/P	tested for 1 minute		4.6kVAC
Insulation Grade				reinforced
Leakage Current				100μA max.
Means of Protection	280VAC working voltage			2MOPP
Medical Device Classification				Type BF
	Notes:			
	Note3: For repeat	t Hi_Pot tacting raduce the	e time and/or the test voltage	

Condition		Value
natural convection 0.1m/s	without derating with derating	-20°C to +50°C -20°C to +80°C
		+85°C
		5000m
non-condensing		95% RH max.
		PD2
according to MIL-HDBK-	+25°C	563 x 10 <sup>3</sup> hours
217F, G.B.	+50°C	112 x 10 <sup>3</sup> hours
		130 x 10 <sup>3</sup> hours
	natural convection 0.1m/s  non-co	natural convection 0.1m/s without derating with derating  non-condensing  according to MIL-HDBK- +25°C



**Series** 

## Specifications (measured @ ta= 25°C, nom. Vin (115/230VAC), full load after warm-up unless otherwise stated)



**Ambient Temperature [°C]** 

Certificate Type (Safety)	Report / File Number	Standard
Information Technology Equipment, General Requirements for Safety (CB Scheme)	T223-0257/17	IEC60950-1:2005, 2nd Edition + Am2:2013 EN60950-1:2006 + A2:2013
Information Technology Equipment, General Requirements for Safety	T223-0257/17	UL60950-1, 2nd Edition:2014 CAN/CSA C22.2 No. 60950-1, 2nd Edition:2014
Medical Electric Equipment, General Requirements for Safety and Essential Performance (CB Scheme)	T223-0256/17	IEC60601-1:2005, AM1:2012 EN60601-1:2006 + A12:2014
Medical Electric Equipment, General Requirements for Safety and Essential Performance	T223-0256/17	CAN/CSA-C22.2 No. 60601-1:14, 3rd Edition 2014 ANSI/AAMI ES60601-1:2005
Household and similar electrical appliances - Safety Part 1: General requirements (CB Scheme)	T211-0761/17	IEC60335-1:2010, 5th Edition + A1:2013 EN60335-1:2012 + A11:2014
Safety of power transformers, power supplies, reactors and similar products for supply voltages up to 1100V		IEC61558-1:2005, 2nd Edition + A1:2009 EN61558-1:2005 + A1:2009
Safety of transformers, reactors, power supply units and similar products for supply voltages up to 1100 V - Part 2-16: Particular requirements and tests for switch mode power supply units	T211-0762/17	IEC61558-2-16:2009, 1st Edition + A1:2013 EN61558-2-16:2009 + A1:2013
EAC	RU-AT.49.09571	TP TC 004/2011 TP TC 004/2011
RoHs 2 (2+)		RoHs 10/10, AM2015
EMC Compliance (Medical)	Condition	Standard / Criterion
Medical electrical equipment Part 1-2: Electromagnetic disturbances – Requirements and tests		EN60601-1-2:2015
ESD Electrostatic discharge immunity test	Air ±2, 4, 8, 15kV; Contact ±8kV	IEC61000-4-2:2008, Criteria A
Radiated, radio-frequency, electromagnetic field immunity test	10V/m (80 - 2700MHz)	IEC61000-4-3:2006 + A2:2010 , Criteria A
Radiated, radio-frequency, electromagnetic field immunity test (table 9)	27V/m (385MHz), 28V/m (450MHz), 9V/m (710, 745, 780MHz), 28V/m (810, 870, 930, 1720, 1845, 1970, 2450MHz), 9V/m (5240, 5500, 5785MHz)	IEC61000-4-3:2006 + A2:2010, Criteria A
Fast Transient and Burst Immunity	AC Power Port ±2.0kV DC Output Port ±1kV	IEC61000-4-4:2012, Criteria A
Surge Immunity	AC Power Port: L-N ±0.5, 1.0kV	IEC61000-4-5:2005, Criteria A



**Series** 

#### **Specifications** (measured @ ta= 25°C, nom. Vin (115/230VAC), full load after warm-up unless otherwise stated)


EMC Compliance (Medical)	Condition	Standard / Criterion
Immunity to conducted disturbances, induced by radio-frequency fields	AC Power Port 6V DC Output Port 6V	IEC61000-4-6:2013, Criteria A
Power Magnetic Field Immunity	50Hz, 60Hz, 30A/m	IEC61000-4-8:2009, Criteria A
Voltage Dips and Interruptions		IEC61000-4-11:2004, Criteria A
EMC Compliance (Household)	Condition	Standard / Criterion
Electromagnetic compatibility of multimedia equipment – Emission Requirements		EN55014-1:2006 + A2:2011
Information technology equipment - Immunity characteristics - Limits and methods of measurement		EN55014-2:1997 + A2:2008
ESD Electrostatic discharge immunity test	Air ±8kV; Contact ±4kV	EN61000-4-2:1995 + A2:2001, Criteria A
Radiated, radio-frequency, electromagnetic field immunity test	3V/m (80 - 1000MHz)	EN61000-4-3:2006 + A1:2008, Criteria A
Fast Transient and Burst Immunity	AC Power Port: ±1.0kV DC Power Port ±0.5kV	EN61000-4-4:2004, Criteria A
Surge Immunity	AC Power Port: L-N ±0.5, 1.0kV	EN61000-4-5:2006, Criteria A
Immunity to conducted disturbances, induced by radio-frequency fields	AC Power Port 3V DC Power Port 3V	EN61000-4-6:2007, Criteria A
Voltage Dips and Interruptions		EN61000-4-11:2004
EMC Compliance (Multimedia)	Condition	Standard / Criterion
Electromagnetic compatibility of multimedia equipment – Emission Requirements		EN55032:2010, Class B
Information technology equipment - Immunity characteristics - Limits and methods of measurement		EN55024:2010
Electromagnetic compatibility of multimedia equipment - Emission requirements		CISPR 32:2012, Class B
ESD Electrostatic discharge immunity test	Air ±2, 4, 8kV; Contact ±4kV	EN61000-4-2:2009, Criteria A
Radiated, radio-frequency, electromagnetic field immunity test	3V/m (80 - 1000MHz)	EN61000-4-3:2006 + A2:2010, Criteria A
Fast Transient and Burst Immunity	AC Power Port: ±1.0kV DC Power Port ±0.5kV	EN61000-4-4:2004, Criteria A
Surge Immunity	AC Power Port: L-N ±0.5, 1.0kV	EN61000-4-5:2006, Criteria A
Immunity to conducted disturbances, induced by radio-frequency fields	AC Power Port 3V DC Power Port 3V	EN61000-4-6:2009, Criteria A
Power Magnetic Field Immunity	50Hz, 60Hz, 1A/m	EN61000-4-8:2010, Criteria A
Voltage Dips and Interruptions		EN61000-4-11:2004
Limits of Voltage Fluctuations & Flicker		IEC/EN61000-3-3:2013

DIMENSION AND PHYSICAL CHARACTERISTICS		
Parameter	Туре	Value
	Case	non-conductive black plastic, (UL94V-0)
Material	Potting	polyurethane, (UL94V-0)
	PCB	FR4, (UL94V-0)
Package Dimension (LxWxH)		53.0 x 51.0 x 24.5mm
Package Weight		88g max.
	continued on next page	



**Series** 

### Specifications (measured @ ta= 25°C, nom. Vin (115/230VAC), full load after warm-up unless otherwise stated)



PACKAGING INFORMATION			
Parameter	Туре	Value	
Packaging Dimension (LxWxH)	carton	310.0 x 220.0 x 100.0mm	
Packaging Quantity		10pcs	
Storage Temperature Range		-30°C to +80°C	
Storage Humidtiy	non-condensing	95% RH max.	

The product information and specifications may be subject to changes even without prior written notice. The product has been designed for various applications; its suitability lies in the responsibility of each customer. The products are not authorized for use in safety-critical applications without RECOM's explicit written consent. A safety-critical application is an application where a failure may reasonably be expected to endanger or cause loss of life, inflict bodily harm or damage property. The applicant shall indemnify and hold harmless RECOM, its affiliated companies and its representatives against any damage claims in connection with the unauthorized use of RECOM products in such safety-critical applications.